



# ZEISS CALYPSO

6.4.08

Part name **Contura G2 10-12-6 1**

Drawing number

Order number

Variant

Company

Department

CMM Type **CONT\_G2**

CMM No. **000000**

Operator **Master**

Text

Last 1 measurements

► Approval ≠ Blocked

Part id **3**




















Time/Date **6/29/2018 1:21 PM**

Run **Current Selection**

No. measured values **10**

No. values: red **0**

Measurement Duration **00:00:01.0**

Name	Value	Nominal Value	+Tol	-Tol	Deviation	+/-
 Radius Measurement1^1	0.4733	0.4724	0.0039	-0.0039	0.0008	
 Radius Measurement1^2	0.4762	0.4724	0.0039	-0.0039	0.0038	
 Radius Measurement1^3	0.4763	0.4724	0.0039	-0.0039	0.0039	
 Radius Measurement1^4	0.4756	0.4724	0.0039	-0.0039	0.0031	
 Radius Measurement1^5	0.4726	0.4724	0.0039	-0.0039	0.0002	
 Radius Measurement1^6	0.4748	0.4724	0.0039	-0.0039	0.0023	
 Radius Measurement1^7	0.4734	0.4724	0.0039	-0.0039	0.0010	
 Radius Measurement1^8	0.4703	0.4724	0.0039	-0.0039	-0.0021	
 Radius Measurement1^9	0.4705	0.4724	0.0039	-0.0039	-0.0020	
 Radius Measurement1^10	0.4736	0.4724	0.0039	-0.0039	0.0012	